



# *Austeofix Surgical Pvt. Ltd.*

## *A Perfect Fixation*

AN ISO 9001:2015, ISO 13485:2016, CE & GMP CERTIFIED COMPANY



**POWER  
TOOLS SYSTEMS**

## Battery Operated Drill Cannulated



### Specification

Diameter of Drill Chuck	0-8mm
Input Voltage	110v-240V/50Hz
Output Power	>120 W
Battery Capacity	1800mA/h
Rated Speed	0-1200 rpm
Rating Torque	6 Nm
Charging Voltage	14.4V
Sterilizing Temperature	135°C -137°C
Noise	<60 db

### Standard Configuration

Handpiece	1
Aseptic Battery Transfer Ring	1
Charger	1
Battery	2
Aluminum Case	1
Drilling Chuck	1
Key	1
Sterilization Case	1



**Sterilization Case**



**Drilling Chuck**

Diameter of Drill Chuck 0-8mm

## Battery Operated Saw



### Specification

Input Voltage	110v-240V/50Hz
Output Power	>120 W
Battery Capacity	1800mA/h
Rated Speed	0-2000 rpm
Charging Voltage	14.4V
Sterilization	HTHP
Sterilizing Temperature	135°C -137°C
Noise	<60 db

### Standard Configuration

Handpiece	1
Aseptic Battery Transfer Ring	1
Charger	1
Battery	2
Aluminum Case	1
Saw Blades	3
Sterilization Case	1



**Sterilization Case**



**Saw Blades**

## Other Accessories



Charger



Battery



® Aseptic Battery Transfer Ring

## Electric Plaster Cutter



### Application & Specification

Cut all kinds of Bandages
Unique Construction
Power Voltage 220V/50Hz
Input Power 115VA

### Standard Configuration

Plaster Cutter	1
Blade	1
Spanner	2
Box For Plaster Cutter	1