



Austeofix Surgical Pvt. Ltd.

A Perfect Fixation

AN ISO 9001:2015, ISO 13485:2016, CE & GMP CERTIFIED COMPANY



***POWER
TOOLS SYSTEMS***



Battery Operated Drill Cannulated

Specification

Diameter of Drill Chuck	0-8mm
Input Voltage	110v-240V/50Hz
Output Power	>120 W
Battery Capacity	1800mA/h
Rated Speed	0-1200 rpm
Rating Torque	6 Nm
Charging Voltage	14.4V
Sterilizing Temperature	135°C -137°C
Noise	<60 db

Standard Configuration

Handpiece	1
Aseptic Battery Transfer Ring	1
Charger	1
Battery	2
Aluminum Case	1
Drilling Chuck	1
Key	1
Sterilization Case	1



Sterilization Case

Drilling Chuck

Diameter of Drill Chuck 0-8mm

Battery Operated Saw

Specification

Input Voltage	110v-240V/50Hz
Output Power	>120 W
Battery Capacity	1800mA/h
Oscillations	16000 CPM
Charging Voltage	14.4V
Sterilization	HTHP
Sterilizing Temperature	135°C -137°C
Noise	<60 db

Standard Configuration

Handpiece	1
Aseptic Battery Transfer Ring	1
Charger	1
Battery	2
Aluminum Case	1
Saw Blades	3
Sterilization Case	1



Saw Blades



Charger



Battery

Upgradable: Above system can be used on electricity using an electric console.